



ASTM E1695-95 Resolution Test for CT-System

date 20/05/08, system SEM- μ CT, s/n 001, voltage 30 kV, current 290 nA



SPREAD:
width=2.0pixels
width=0.9microns
MTF (FOURIER):
N for 10%=323
resolution:
1.6 pixels
0.7 microns